
Midmark M11 - v1.0.0
Total Cycles: 3745

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2023 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.8 F 0.0
2:00 167.6 F 0.0
4:00 184.9 F 2.7
6:00 201.3 F 6.8
8:00 217.3 F 11.2
10:00 223.3 F 5.7
12:00 235.9 F 11.0
14:00 234.0 F 8.3
16:00 241.4 F 10.9
18:00 246.9 F 13.6
20:00 258.9 F 20.1
22:00 269.8 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.1 F 28.5
2:00 272.6 F 28.9
3:00 271.6 F 28.5
4:00 272.0 F 28.5

Min 271.5 F 28.2
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 209.6 F 0.0
10:00 231.8 F 0.0
15:00 227.3 F 0.0
20:00 220.1 F 0.0
25:00 216.5 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 22:53
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:02:26
CYCLE COMPLETE

09/08/2023 04:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____