
Midmark M11 - v1.0.0
Total Cycles: 2537

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 08:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.4 F 0.0
2:00 82.1 F 0.0
4:00 92.2 F 0.0
6:00 125.7 F 0.1
8:00 153.0 F 2.6
10:00 178.4 F 5.9
12:00 197.4 F 9.4
14:00 206.9 F 12.9
16:00 216.1 F 5.6
18:00 226.9 F 9.5
20:00 237.3 F 11.7
22:00 236.1 F 10.2
24:00 236.8 F 8.3
26:00 247.0 F 13.9
28:00 258.0 F 19.9
30:00 268.3 F 26.2
32:00 272.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 28.9
1:00 272.7 F 29.6
2:00 271.9 F 28.8
3:00 272.8 F 29.5
4:00 271.8 F 28.8

Min 271.7 F 28.8
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 204.3 F 0.0
10:00 211.0 F 0.0
15:00 219.3 F 0.0
20:00 222.2 F 0.0
25:00 232.5 F 0.0
30:00 227.5 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:01
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

11/10/2022 09:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____