
Midmark M11 - v1.0.5
Total Cycles: 471

Name: RITTER2
Sterilizer ID: V2313228
10/22/2021 06:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 73.4 F 0.0
4:00 74.9 F 0.0
6:00 96.4 F 0.1
8:00 139.2 F 0.3
10:00 173.9 F 1.3
12:00 187.7 F 4.3
14:00 196.0 F 7.8
16:00 211.3 F 11.5
18:00 226.0 F 6.2
20:00 230.7 F 11.1
22:00 233.8 F 8.2
24:00 243.2 F 11.8
26:00 244.4 F 12.9
28:00 255.5 F 19.0
30:00 265.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.3
1:00 272.7 F 30.4
2:00 271.8 F 29.6
3:00 272.7 F 30.2
4:00 271.8 F 29.9

Min 270.9 F 29.1
Max 273.3 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 224.0 F 0.0
10:00 214.1 F 0.0
15:00 216.0 F 0.0
20:00 215.9 F 0.0
25:00 213.5 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:09:54
CYCLE COMPLETE

10/22/2021 07:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____