
Midmark M11 - v1.0.5
Total Cycles: 2712

Name: RITTER2
Sterilizer ID: V2313228
07/31/2023 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.0 F 0.0
2:00 90.9 F 0.0
4:00 107.8 F 0.0
6:00 147.7 F 0.0
8:00 175.6 F 0.0
10:00 197.5 F 0.0
12:00 208.7 F 0.0
14:00 217.3 F 2.4
16:00 226.0 F 5.4
18:00 234.6 F 8.9
20:00 243.7 F 13.1
22:00 241.1 F 11.2
24:00 238.7 F 9.9
26:00 238.9 F 9.5
28:00 250.7 F 15.6
30:00 261.7 F 22.0
32:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.7 F 29.0
2:00 271.7 F 28.5
3:00 272.7 F 29.0
4:00 271.7 F 28.4

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 202.6 F 0.0
10:00 210.4 F 0.0
15:00 212.4 F 0.0
20:00 214.0 F 0.0
25:00 212.6 F 0.0
30:00 210.2 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:34
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:10:54
CYCLE COMPLETE

07/31/2023 02:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____