
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000943
12:25 18/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:24
STERILIZING 4:34
136.0 C 228kPa 4:34
136.0 C 229kPa 5:04
136.2 C 231kPa 5:34
136.2 C 231kPa 6:04
136.4 C 233kPa 6:34
136.2 C 231kPa 7:04
136.4 C 231kPa 7:34
136.4 C 233kPa 8:04
136.4 C 233kPa 8:04
VENTING 8:05
AIR DRYING 9:01
CYCLE COMPLETE 24:01

Digital Signature #
CD1C57EEC8ECE8A5
