
Midmark M11 - v1.0.5
Total Cycles: 3344

Name: RITTER2
Sterilizer ID: V2313228
02/01/2024 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 155.2 F 1.0
4:00 169.6 F 3.9
6:00 190.5 F 8.3
8:00 209.2 F 13.0
10:00 225.3 F 8.2
12:00 239.7 F 13.5
14:00 238.4 F 10.8
16:00 236.8 F 9.4
18:00 250.6 F 16.0
20:00 262.1 F 22.6
22:00 272.5 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 271.7 F 29.3
2:00 272.3 F 29.6
3:00 272.3 F 29.9
4:00 271.0 F 29.4

Min 271.0 F 29.3
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 205.4 F 0.0
10:00 196.1 F 0.0
15:00 195.1 F 0.0
20:00 192.9 F 0.0
25:00 192.0 F 0.0
30:00 189.8 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 23:08
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

02/01/2024 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____