

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3157

Name: RITTER2  
Sterilizer ID: V2313228  
11/23/2023 10:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 70.8 F 0.0  
4:00 76.4 F 0.2  
6:00 104.0 F 1.4  
8:00 131.2 F 4.0  
10:00 163.6 F 7.5  
12:00 183.1 F 11.4  
14:00 207.3 F 3.8  
16:00 219.0 F 7.7  
18:00 226.8 F 11.9  
20:00 229.0 F 7.9  
22:00 241.3 F 13.1  
24:00 240.2 F 11.4  
26:00 252.6 F 17.5  
28:00 263.7 F 24.0  
30:00 271.8 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 271.8 F 29.8  
2:00 272.3 F 29.6  
3:00 271.3 F 29.4  
4:00 272.4 F 30.0

Min 271.1 F 29.0  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 202.1 F 0.0  
10:00 203.5 F 0.0  
15:00 189.8 F 0.0  
20:00 188.0 F 0.0  
25:00 186.1 F 0.0  
30:00 185.2 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 30:02  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:08:17  
CYCLE COMPLETE

11/23/2023 11:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_