
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001885
9:12 24/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 2:03
STERILIZING 3:03
136.4 C 224kPa 3:03
136.3 C 224kPa 3:33
136.4 C 225kPa 4:03
136.0 C 222kPa 4:33
136.4 C 225kPa 5:03
136.4 C 222kPa 5:33
136.4 C 226kPa 6:03
136.4 C 225kPa 6:33
136.4 C 225kPa 6:33
VENTING 6:33
AIR DRYING 7:26
CYCLE COMPLETE 37:26

Digital Signature #
1C51111FB7A2A470
