

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1720

Name: RITTER2  
Sterilizer ID: V2313228  
11/14/2022 04:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
2:00 189.5 F 3.2  
4:00 206.3 F 9.0  
6:00 224.5 F 8.4  
8:00 235.2 F 11.7  
10:00 237.4 F 10.2  
12:00 237.5 F 9.4  
14:00 252.6 F 17.0  
16:00 265.2 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.7 F 29.5  
2:00 271.5 F 28.7  
3:00 271.8 F 29.4  
4:00 273.3 F 29.8  
  
Min 271.0 F 28.6  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
5:00 221.1 F 0.0  
10:00 233.1 F 0.0  
15:00 227.4 F 0.0  
20:00 226.4 F 0.0  
25:00 225.3 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 17:44  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:56:59  
CYCLE COMPLETE

11/14/2022 05:37 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_