
Midmark M11 - v1.0.0
Total Cycles: 2521

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2022 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 177.8 F 3.7
4:00 191.6 F 8.1
6:00 212.8 F 13.2
8:00 226.6 F 7.5
10:00 233.8 F 12.8
12:00 236.0 F 9.8
14:00 235.3 F 8.3
16:00 248.2 F 14.5
18:00 260.0 F 21.1
20:00 271.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 29.1
2:00 272.0 F 29.2
3:00 272.0 F 28.8
4:00 272.4 F 29.4

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 216.3 F 0.0
10:00 233.8 F 0.0
15:00 236.7 F 0.0
20:00 243.0 F 0.0
25:00 228.6 F 0.0
30:00 220.7 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:00
CYCLE COMPLETE

11/07/2022 04:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____