
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751

SN 500919I00001

Unit # : 000

WATER QUALITY

0.2uS / 0.1ppm

CYCLE NUMBER 005037

10:11 01/02/2023

UNWRAPPED

135 C for 3.5 min.

CYCLE START 0:00

CONDITIONING 2:37

PRESSURIZING 2:38

STERILIZING 3:33

136.4 C 223kPa 3:33

136.0 C 220kPa 4:03

136.2 C 221kPa 4:33

136.4 C 224kPa 5:03

136.3 C 221kPa 5:33

136.4 C 223kPa 6:03

136.3 C 221kPa 6:33

136.4 C 223kPa 7:03

136.4 C 223kPa 7:03

VENTING 7:03

AIR DRYING 7:53

CYCLE COMPLETE 11:15

Digital Signature #

92580C3ACE978067
