

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 003704  
9:47 27/05/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:17  
STERILIZING 3:14  
136.0 C 220kPa 3:14  
136.4 C 222kPa 3:44  
136.4 C 224kPa 4:14  
136.4 C 223kPa 4:44  
136.3 C 222kPa 5:14  
136.3 C 220kPa 5:44  
136.4 C 222kPa 6:14  
136.4 C 223kPa 6:44  
136.4 C 223kPa 6:44  
VENTING 6:44  
AIR DRYING 7:31  
CYCLE COMPLETE 7:45

-----  
Digital Signature #  
2E80AD23CDE90AB2  
-----