
Midmark M11 - v1.0.0
Total Cycles: 2576

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2022 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 194.6 F 4.8
4:00 207.5 F 9.6
6:00 219.6 F 6.3
8:00 225.3 F 9.2
10:00 234.4 F 9.4
12:00 242.0 F 12.0
14:00 240.0 F 10.5
16:00 253.6 F 17.3
18:00 265.3 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 271.2 F 28.6
2:00 272.2 F 29.4
3:00 272.4 F 29.3
4:00 272.3 F 29.0

Min 271.0 F 28.3
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 222.5 F 0.0
10:00 234.5 F 0.0
15:00 232.8 F 0.0
20:00 227.3 F 0.0
25:00 226.7 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

11/17/2022 01:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____