
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002019
9:25 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:06
STERILIZING 3:12
136.0 C 221kPa 3:12
136.4 C 226kPa 3:42
136.4 C 224kPa 4:12
136.4 C 225kPa 4:42
136.4 C 226kPa 5:12
136.2 C 223kPa 5:42
136.3 C 224kPa 6:12
136.4 C 224kPa 6:42
136.4 C 224kPa 6:42
VENTING 6:42
AIR DRYING 7:34
CYCLE COMPLETE 37:34

Digital Signature #
005A6F7717EE043B
