

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001750  
11:21 14/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:15  
STERILIZING 4:24  
136.5 C 222kPa 4:24  
136.4 C 224kPa 4:54  
136.3 C 224kPa 5:24  
136.4 C 225kPa 5:54  
136.3 C 223kPa 6:24  
136.4 C 224kPa 6:54  
136.1 C 219kPa 7:24  
136.3 C 222kPa 7:54  
136.3 C 222kPa 7:54  
VENTING 7:54  
AIR DRYING 8:44  
CYCLE COMPLETE 8:47

-----  
Digital Signature #  
C7CF5FAE65530DC7  
-----