

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1020

Name: RITTER2  
Sterilizer ID: V2313228  
04/22/2022 09:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.1 F 0.0  
2:00 86.2 F 0.0  
4:00 128.7 F 0.0  
6:00 138.9 F 0.0  
8:00 172.7 F 0.3  
10:00 202.1 F 1.7  
12:00 216.5 F 5.7  
14:00 227.6 F 9.6  
16:00 234.8 F 10.3  
18:00 238.8 F 10.6  
20:00 235.0 F 8.2  
22:00 240.6 F 10.0  
24:00 249.2 F 14.7  
26:00 262.5 F 22.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.3  
1:00 272.3 F 28.9  
2:00 272.8 F 28.9  
3:00 272.4 F 28.5  
4:00 272.0 F 28.2

Min 271.7 F 28.1  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 220.9 F 0.0  
5:00 227.3 F 0.0  
10:00 221.0 F 0.0  
15:00 215.6 F 0.0  
20:00 209.8 F 0.0  
Dry Cycle Aborted

FILLING: 1:12  
HEATING: 27:59  
STERILIZING: 4:00  
VENTING: 2:32  
DRYING: 20:06  
TOTAL CYCLE: 00:55:49

04/22/2022 10:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_