
Midmark M11 - v1.0.5
Total Cycles: 517

Name: RITTER2
Sterilizer ID: V2313228
11/03/2021 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.6 F 0.0
2:00 140.2 F 1.6
4:00 147.0 F 4.4
6:00 165.1 F 8.2
8:00 185.4 F 12.9
10:00 208.8 F 6.8
12:00 224.2 F 11.4
14:00 227.4 F 8.2
16:00 240.8 F 12.5
18:00 242.4 F 12.5
20:00 254.6 F 18.7
22:00 265.4 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.4 F 30.3
2:00 272.1 F 29.8
3:00 271.3 F 29.3
4:00 271.8 F 30.1

Min 271.2 F 29.3
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 201.2 F 0.0
10:00 208.1 F 0.0
15:00 198.6 F 0.0
20:00 191.4 F 0.0
25:00 209.0 F 0.0
30:00 198.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

11/03/2021 01:16 PM
ACCEPTED: X
REJECTED:
Operator: U01