
Midmark M11 - v1.0.0
Total Cycles: 4072

Name: RITTER 1
Sterilizer ID: V2178198
12/01/2023 09:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	69.2 F	0.0
2:00	70.0 F	0.2
4:00	75.5 F	1.0
6:00	100.5 F	2.8
8:00	129.1 F	5.1
10:00	154.8 F	7.7
12:00	175.6 F	10.7
14:00	201.5 F	9.2
16:00	204.6 F	4.0
18:00	214.5 F	6.9
20:00	224.2 F	10.2
22:00	234.8 F	10.6
24:00	234.9 F	9.5
26:00	240.3 F	10.4
28:00	243.8 F	12.4
30:00	255.0 F	18.0
32:00	264.5 F	23.7
34:00	272.2 F	28.7

STERILIZING

mm:ss	Degrees	PSI
0:00	271.2 F	28.4
1:00	272.4 F	29.2
2:00	272.1 F	28.8
3:00	271.1 F	28.3
4:00	272.2 F	29.2

Min 271.0 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
0:00	187.2 F	0.0	
5:00	214.7 F	0.0	
10:00	213.7 F	0.0	
15:00	216.1 F	0.0	
20:00	214.5 F	0.0	
25:00	219.7 F	0.0	
30:00	222.0 F	0.0	

Drying Complete

FILLING: 0:52
HEATING: 34:15
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:12:53
CYCLE COMPLETE

12/01/2023 10:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____