

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002283  
12:52 18/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:17  
PRESSURIZING 2:20  
STERILIZING 3:28  
136.5 C 227kPa 3:28  
136.2 C 225kPa 3:59  
136.4 C 227kPa 4:29  
136.4 C 227kPa 4:59  
136.4 C 225kPa 5:29  
136.3 C 226kPa 5:59  
136.1 C 224kPa 6:29  
136.1 C 224kPa 6:59  
136.1 C 224kPa 6:59  
VENTING 6:59  
AIR DRYING 7:55  
CYCLE COMPLETE 16:50  
-----

Digital Signature #  
272CD7A2AC9B3EF3  
-----