
Midmark M11 - v1.0.5
Total Cycles: 2707

Name: RITTER2
Sterilizer ID: V2313228
07/27/2023 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 74.7 F 0.0
4:00 103.0 F 0.0
6:00 97.5 F 0.0
8:00 140.7 F 0.0
10:00 183.6 F 0.4
12:00 191.6 F 2.9
14:00 212.6 F 5.8
16:00 220.0 F 9.3
18:00 232.5 F 13.1
20:00 232.2 F 9.1
22:00 239.1 F 11.4
24:00 242.4 F 12.1
26:00 240.7 F 11.0
28:00 253.4 F 17.3
30:00 264.3 F 23.8
32:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.3 F 29.1
2:00 271.9 F 29.2
3:00 271.9 F 28.7
4:00 272.5 F 29.5

Min 271.4 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 212.6 F 0.0
10:00 208.3 F 0.0
15:00 209.9 F 0.0
20:00 216.6 F 0.0
25:00 231.3 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 32:05
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:07
CYCLE COMPLETE

07/27/2023 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____