

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3341

Name: RITTER 1  
Sterilizer ID: V2178198  
06/06/2023 10:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
2:00 183.9 F 0.0  
4:00 201.8 F 3.9  
6:00 222.3 F 8.8  
8:00 227.9 F 10.1  
10:00 234.5 F 10.1  
12:00 235.1 F 8.4  
14:00 245.6 F 12.9  
16:00 244.6 F 12.5  
18:00 257.3 F 19.1  
20:00 268.4 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.7 F 28.9  
2:00 271.6 F 28.1  
3:00 272.5 F 29.0  
4:00 272.2 F 28.5  
  
Min 271.4 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 212.8 F 0.0  
10:00 225.2 F 0.0  
15:00 227.9 F 0.0  
20:00 226.8 F 0.0  
25:00 220.7 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 21:15  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

06/06/2023 11:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_