
Midmark M11 - v1.0.5
Total Cycles: 257

Name: RITTER2
Sterilizer ID: V2313228
07/28/2021 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 171.4 F 3.1
4:00 185.5 F 7.5
6:00 203.2 F 12.9
8:00 222.0 F 7.9
10:00 235.1 F 13.5
12:00 238.2 F 11.2
14:00 238.3 F 10.2
16:00 251.7 F 16.9
18:00 263.3 F 23.7
20:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.3
1:00 272.4 F 30.1
2:00 272.1 F 29.3
3:00 272.0 F 29.9
4:00 272.5 F 29.8

Min 271.4 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 210.1 F 0.0
10:00 211.7 F 0.0
15:00 212.3 F 0.0
20:00 206.3 F 0.0
25:00 197.2 F 0.0
Dry Cycle Aborted

FILLING: 1:38
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:09
DRYING: 29:24
TOTAL CYCLE: 00:58:13

07/28/2021 10:57 AM
ACCEPTED: X
REJECTED:
Operator: U01