
Midmark M11 - v1.0.0
Total Cycles: 5225

Name: RITTER 1
Sterilizer ID: V2178198
01/08/2025 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.4 F 0.0
2:00 74.2 F 0.0
4:00 84.6 F 0.0
6:00 115.9 F 0.0
8:00 148.3 F 0.2
10:00 179.2 F 1.0
12:00 197.5 F 4.4
14:00 207.4 F 7.6
16:00 218.9 F 11.0
18:00 223.8 F 5.8
20:00 232.3 F 9.0
22:00 242.7 F 13.5
24:00 240.9 F 10.9
26:00 237.2 F 9.0
28:00 251.2 F 15.5
30:00 262.5 F 22.0
32:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 271.5 F 27.9
2:00 271.6 F 28.1
3:00 271.6 F 28.1
4:00 271.6 F 28.1

Min 271.5 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 192.1 F 0.0
10:00 194.9 F 0.0
15:00 197.5 F 0.0
20:00 205.4 F 0.0
25:00 222.9 F 0.0
30:00 226.3 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 32:15
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:27
CYCLE COMPLETE

01/08/2025 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____