
Midmark M11 - v1.0.0
Total Cycles: 4470

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2024 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.4 F 0.2
4:00 77.9 F 1.0
6:00 101.0 F 3.0
8:00 141.6 F 5.4
10:00 168.1 F 8.2
12:00 192.3 F 11.2
14:00 215.9 F 4.5
16:00 214.4 F 4.6
18:00 221.3 F 7.5
20:00 231.6 F 10.9
22:00 231.2 F 7.6
24:00 239.7 F 10.6
26:00 235.5 F 8.3
28:00 246.8 F 13.5
30:00 257.1 F 19.0
32:00 266.5 F 24.9
34:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 28.7
2:00 272.7 F 29.1
3:00 271.2 F 28.5
4:00 272.7 F 28.9

Min 271.1 F 28.4
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.9 F 0.0
5:00 198.6 F 0.0
10:00 183.1 F 0.0
15:00 179.6 F 0.0
20:00 178.6 F 0.0
25:00 180.2 F 0.0
30:00 184.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:27
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:13:02
CYCLE COMPLETE

04/18/2024 10:08 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____