
Midmark M11 - v1.0.0
Total Cycles: 4159

Name: RITTER 1
Sterilizer ID: V2178198
12/27/2023 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 125.4 F 1.7
4:00 143.7 F 4.5
6:00 183.4 F 8.1
8:00 198.9 F 13.3
10:00 213.7 F 5.5
12:00 228.9 F 9.9
14:00 235.3 F 8.7
16:00 239.6 F 11.3
18:00 237.4 F 9.2
20:00 250.7 F 15.7
22:00 262.2 F 22.2
24:00 272.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.9 F 28.4
3:00 272.1 F 29.0
4:00 272.5 F 29.0

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 216.8 F 0.0
10:00 221.0 F 0.0
15:00 221.4 F 0.0
20:00 223.5 F 0.0
25:00 218.1 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:24
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:02:48
CYCLE COMPLETE

12/27/2023 03:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____