
Midmark M11 - v1.0.0
Total Cycles: 5246

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2025 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.3 F 0.0
2:00 94.2 F 0.0
4:00 120.6 F 0.0
6:00 149.6 F 0.1
8:00 178.4 F 0.3
10:00 197.3 F 3.1
12:00 209.6 F 6.3
14:00 223.0 F 9.6
16:00 232.8 F 12.7
18:00 233.1 F 8.7
20:00 244.5 F 13.4
22:00 241.7 F 11.0
24:00 237.1 F 9.0
26:00 250.7 F 15.3
28:00 261.6 F 21.4
30:00 271.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.1 F 28.7
2:00 272.4 F 28.5
3:00 271.5 F 27.9
4:00 271.5 F 28.3

Min 271.1 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 202.6 F 0.0
10:00 194.8 F 0.0
15:00 228.2 F 0.0
20:00 235.9 F 0.0
25:00 232.3 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:45
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

01/15/2025 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____