
Midmark M11 - v1.0.0
Total Cycles: 2022

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2022 07:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 86.0 F 0.0
4:00 87.0 F 0.2
6:00 106.4 F 1.3
8:00 142.2 F 2.9
10:00 167.7 F 5.1
12:00 187.2 F 8.1
14:00 205.1 F 11.7
16:00 217.0 F 3.8
18:00 225.1 F 7.8
20:00 235.3 F 11.9
22:00 233.3 F 7.8
24:00 245.1 F 13.4
26:00 242.9 F 11.6
28:00 254.8 F 17.8
30:00 265.3 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.6
2:00 272.4 F 29.0
3:00 271.5 F 28.1
4:00 272.4 F 28.8

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 230.3 F 0.0
10:00 223.3 F 0.0
15:00 217.1 F 0.0
20:00 232.7 F 0.0
25:00 237.3 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:52
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:47
CYCLE COMPLETE

06/15/2022 08:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____