
Midmark M11 - v1.0.0
Total Cycles: 5358

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2025 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.8 F 0.0
2:00 168.6 F 2.2
4:00 169.1 F 5.5
6:00 198.2 F 9.3
8:00 224.6 F 7.5
10:00 219.9 F 6.6
12:00 231.9 F 10.8
14:00 232.6 F 7.6
16:00 242.4 F 12.5
18:00 240.8 F 10.6
20:00 252.7 F 16.4
22:00 263.1 F 22.5
24:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.4
2:00 272.4 F 29.2
3:00 272.0 F 28.8
4:00 271.2 F 28.6

Min 271.2 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 231.0 F 0.0
10:00 204.3 F 0.0
15:00 188.9 F 0.0
20:00 185.5 F 0.0
25:00 182.0 F 0.0
30:00 182.6 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 25:20
STERILIZING: 4:00
VENTING: 4:00
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

02/24/2025 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____