
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002369
9:10 10/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:11
STERILIZING 4:21
136.5 C 226kPa 4:21
136.1 C 221kPa 4:51
136.3 C 222kPa 5:21
136.5 C 222kPa 5:51
136.3 C 224kPa 6:21
136.3 C 224kPa 6:51
136.1 C 222kPa 7:21
136.4 C 222kPa 7:51
136.4 C 222kPa 7:51
VENTING 7:51
AIR DRYING 8:40
CYCLE COMPLETE 26:38

Digital Signature #
1CF4E4E4BE1EB9B0
