
Midmark M11 - v1.0.5
Total Cycles: 2650

Name: RITTER2
Sterilizer ID: V2313228
07/14/2023 04:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.9 F 0.0
2:00 160.5 F 1.9
4:00 173.4 F 5.6
6:00 201.4 F 10.6
8:00 217.0 F 5.2
10:00 229.3 F 10.7
12:00 231.0 F 7.7
14:00 242.1 F 11.9
16:00 244.8 F 13.1
18:00 257.4 F 19.9
20:00 268.7 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.7
2:00 271.5 F 28.9
3:00 272.5 F 29.8
4:00 272.1 F 29.2

Min 271.3 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 225.0 F 0.0
10:00 230.3 F 0.0
15:00 225.1 F 0.0
20:00 223.0 F 0.0
25:00 221.6 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

07/14/2023 05:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____