

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1512

Name: RITTER 1  
Sterilizer ID: V2178198  
01/12/2022 08:08 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.4 F 0.0  
2:00 76.4 F 0.0  
4:00 87.8 F 0.0  
6:00 115.8 F 1.7  
8:00 145.0 F 4.2  
10:00 166.7 F 7.2  
12:00 185.5 F 10.9  
14:00 208.0 F 3.9  
16:00 216.5 F 6.2  
18:00 226.3 F 9.9  
20:00 235.6 F 10.8  
22:00 237.8 F 10.4  
24:00 235.3 F 8.1  
26:00 247.7 F 14.2  
28:00 259.0 F 20.3  
30:00 269.1 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.1 F 28.7  
2:00 272.5 F 29.2  
3:00 271.5 F 28.6  
4:00 272.0 F 28.7

Min 271.4 F 28.4  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.6 F 0.0  
5:00 216.1 F 0.0  
10:00 228.0 F 0.0  
15:00 220.4 F 0.0  
20:00 223.4 F 0.0  
25:00 218.3 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 31:03  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:09:02  
CYCLE COMPLETE

01/12/2022 09:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_