

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002103  
8:05 02/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:00  
PRESSURIZING 2:08  
STERILIZING 3:13  
136.0 C 220kPa 3:13  
136.4 C 225kPa 3:43  
136.2 C 223kPa 4:13  
136.3 C 223kPa 4:43  
136.2 C 223kPa 5:13  
136.4 C 224kPa 5:43  
136.4 C 224kPa 6:13  
136.3 C 224kPa 6:43  
136.3 C 224kPa 6:43  
VENTING 6:43  
AIR DRYING 7:33  
CYCLE COMPLETE 37:33  
-----

Digital Signature #  
4D8E6DA3D586E143  
-----