
Midmark M11 - v1.0.0
Total Cycles: 2210

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2022 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 191.7 F 4.2
4:00 203.1 F 9.3
6:00 221.0 F 6.7
8:00 231.7 F 10.3
10:00 233.2 F 7.7
12:00 242.6 F 12.0
14:00 245.1 F 13.1
16:00 258.2 F 20.0
18:00 269.7 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.2 F 29.1
2:00 271.3 F 28.5
3:00 272.3 F 29.2
4:00 271.5 F 28.6

Min 271.2 F 28.4
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.6 F 0.0
5:00 215.8 F 0.0
10:00 225.8 F 0.0
15:00 227.7 F 0.0
20:00 229.9 F 0.0
25:00 230.0 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:58:01
CYCLE COMPLETE

08/08/2022 05:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____