
Midmark M11 - v1.0.0
Total Cycles: 1765

Name: RITTER 1
Sterilizer ID: V2178198
03/30/2022 06:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.2 F 0.0
2:00 101.9 F 0.0
4:00 166.1 F 0.0
6:00 184.3 F 0.3
8:00 194.6 F 2.3
10:00 204.5 F 5.1
12:00 216.6 F 8.5
14:00 228.7 F 12.5
16:00 225.1 F 5.7
18:00 235.9 F 10.0
20:00 237.2 F 8.9
22:00 242.2 F 11.5
24:00 238.2 F 9.3
26:00 252.0 F 16.0
28:00 263.2 F 22.5
30:00 272.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 271.7 F 28.5
2:00 272.3 F 28.4
3:00 271.2 F 28.0
4:00 272.5 F 28.8

Min 271.1 F 27.8
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 194.4 F 0.0
10:00 208.8 F 0.0
15:00 226.3 F 0.0
20:00 235.2 F 0.0
25:00 224.9 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:04
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:07:50
CYCLE COMPLETE

03/30/2022 07:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____