
Midmark M11 - v1.0.0
Total Cycles: 4946

Name: RITTER 1
Sterilizer ID: V2178198
10/01/2024 10:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.1 F 0.0
2:00 157.3 F 0.1
4:00 171.2 F 0.4
6:00 193.3 F 1.9
8:00 210.5 F 5.7
10:00 223.7 F 9.6
12:00 230.9 F 10.7
14:00 234.5 F 9.4
16:00 239.4 F 10.3
18:00 244.3 F 12.3
20:00 241.8 F 11.0
22:00 254.5 F 17.4
24:00 265.5 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.5 F 28.7
2:00 272.0 F 28.2
3:00 271.3 F 28.2
4:00 272.4 F 28.8

Min 271.2 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 232.0 F 0.0
10:00 238.3 F 0.0
15:00 231.1 F 0.0
20:00 224.9 F 0.0
25:00 227.1 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 25:50
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:23
CYCLE COMPLETE

10/01/2024 11:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____