
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000254
8:39 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:14
STERILIZING 4:54
136.4 C 232kPa 4:54
136.4 C 233kPa 5:24
136.3 C 232kPa 5:54
136.3 C 232kPa 6:24
136.4 C 233kPa 6:54
136.2 C 231kPa 7:24
136.3 C 233kPa 7:54
136.3 C 232kPa 8:24
136.3 C 232kPa 8:24
VENTING 8:25
AIR DRYING 9:28
CYCLE COMPLETE 24:28

Digital Signature #
F9CE286778CCFAD1
