

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4160

Name: RITTER2  
Sterilizer ID: V2313228  
11/27/2024 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.6 F 0.0  
2:00 73.8 F 0.0  
4:00 78.6 F 0.0  
6:00 107.5 F 0.2  
8:00 153.2 F 0.5  
10:00 181.4 F 0.9  
12:00 199.2 F 3.4  
14:00 211.2 F 6.9  
16:00 222.7 F 10.6  
18:00 228.1 F 7.8  
20:00 237.1 F 10.8  
22:00 234.4 F 8.4  
24:00 241.3 F 10.8  
26:00 247.5 F 14.2  
28:00 259.6 F 20.9  
30:00 270.6 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.4 F 29.0  
2:00 272.4 F 29.3  
3:00 271.9 F 28.6  
4:00 272.8 F 29.4

Min 271.6 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.4 F 0.0  
5:00 204.9 F 0.0  
10:00 191.1 F 0.0  
15:00 190.3 F 0.0  
20:00 190.9 F 0.0  
25:00 203.3 F 0.0  
30:00 210.0 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:46  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:09:03  
CYCLE COMPLETE

11/27/2024 11:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_