

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2894

Name: RITTER 1  
Sterilizer ID: V2178198  
02/13/2023 04:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
2:00 186.0 F 4.5  
4:00 199.4 F 9.4  
6:00 215.4 F 6.9  
8:00 229.9 F 9.1  
10:00 235.6 F 9.9  
12:00 241.3 F 12.1  
14:00 240.2 F 10.8  
16:00 254.0 F 17.7  
18:00 265.8 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.6 F 29.4  
2:00 271.7 F 28.7  
3:00 272.3 F 29.5  
4:00 272.3 F 29.2  
  
Min 271.5 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 220.2 F 0.0  
10:00 232.4 F 0.0  
15:00 231.5 F 0.0  
20:00 224.9 F 0.0  
25:00 219.5 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:41  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

02/13/2023 05:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_