
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008012
10:11 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:47
STERILIZING 4:46
136.0 C 220kPa 4:46
136.4 C 225kPa 5:16
136.3 C 223kPa 5:46
136.0 C 222kPa 6:16
136.3 C 221kPa 6:46
135.9 C 221kPa 7:16
136.5 C 223kPa 7:46
136.4 C 224kPa 8:16
136.4 C 224kPa 8:16
VENTING 8:16
AIR DRYING 9:09
CYCLE COMPLETE 10:15

Digital Signature #
4532324DCBD014F2
