
Midmark M11 - v1.0.0
Total Cycles: 3616

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2023 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 89.1 F 0.0
4:00 93.4 F 0.0
6:00 124.2 F 0.0
8:00 148.8 F 0.3
10:00 174.5 F 1.2
12:00 193.8 F 4.1
14:00 213.3 F 7.1
16:00 221.6 F 10.3
18:00 223.9 F 9.0
20:00 232.2 F 8.2
22:00 242.0 F 12.6
24:00 237.9 F 9.8
26:00 238.7 F 9.3
28:00 248.2 F 14.2
30:00 259.4 F 20.3
32:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.7
2:00 271.6 F 28.0
3:00 272.7 F 28.9
4:00 271.9 F 28.3

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 226.4 F 0.0
10:00 195.6 F 0.0
15:00 194.5 F 0.0
20:00 247.9 F 0.0
25:00 234.8 F 0.0
30:00 226.8 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:11:20
CYCLE COMPLETE

08/04/2023 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____