

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4024

Name: RITTER2  
Sterilizer ID: V2313228  
10/11/2024 12:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.8 F 0.0  
2:00 125.3 F 0.0  
4:00 131.2 F 0.7  
6:00 163.0 F 3.9  
8:00 191.0 F 7.5  
10:00 204.7 F 11.4  
12:00 216.6 F 5.5  
14:00 227.4 F 10.2  
16:00 232.0 F 8.6  
18:00 239.7 F 12.2  
20:00 238.7 F 10.5  
22:00 250.9 F 16.5  
24:00 261.7 F 22.7  
26:00 271.8 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.3 F 30.0  
2:00 272.1 F 29.4  
3:00 271.0 F 29.2  
4:00 272.1 F 30.2

Min 271.0 F 29.0  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 201.5 F 0.0  
10:00 200.2 F 0.0  
15:00 192.3 F 0.0  
20:00 187.9 F 0.0  
25:00 188.6 F 0.0  
30:00 186.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:39  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:05:09  
CYCLE COMPLETE

10/11/2024 01:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_