
Midmark M11 - v1.0.5
Total Cycles: 1580

Name: RITTER2
Sterilizer ID: V2313228
10/11/2022 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	163.7 F	0.0
2:00	161.3 F	0.2
4:00	194.9 F	1.1
6:00	208.6 F	7.2
8:00	227.8 F	12.3
10:00	236.7 F	10.7
12:00	236.7 F	9.2
14:00	237.4 F	8.8
16:00	252.1 F	16.4
18:00	246.3 F	11.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 244F
CHAMBER PRESS 10.7PSI
** ITEMS NOT STERILE **

10/11/2022 11:08 AM