

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001688  
11:36 24/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:02  
PRESSURIZING 2:06  
STERILIZING 3:05  
136.4 C 224kPa 3:05  
136.4 C 226kPa 3:35  
136.4 C 226kPa 4:05  
136.4 C 226kPa 4:35  
136.1 C 224kPa 5:05  
136.4 C 226kPa 5:35  
136.4 C 225kPa 6:05  
136.4 C 226kPa 6:35  
136.4 C 226kPa 6:35  
VENTING 6:35  
AIR DRYING 7:27  
CYCLE COMPLETE 11:02  
-----

Digital Signature #  
A05989319960302A  
-----