
Midmark M11 - v1.0.0
Total Cycles: 775

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2021 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.2 F 0.0
2:00 110.3 F 0.0
4:00 116.2 F 0.1
6:00 152.0 F 2.0
8:00 189.0 F 6.1
10:00 222.0 F 10.7
12:00 224.5 F 5.6
14:00 234.5 F 10.0
16:00 235.4 F 8.2
18:00 242.1 F 12.1
20:00 239.5 F 10.1
22:00 252.4 F 16.5
24:00 263.9 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.0 F 28.4
2:00 272.1 F 28.6
3:00 272.3 F 28.8
4:00 272.5 F 29.0

Min 271.5 F 28.2
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 215.4 F 0.0
10:00 210.5 F 0.0
15:00 202.1 F 0.0
20:00 214.2 F 0.0
25:00 217.5 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:50
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:03:43
CYCLE COMPLETE

04/07/2021 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____