
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006077
10:59 04/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:16
STERILIZING 4:26
136.4 C 222kPa 4:26
136.3 C 221kPa 4:56
136.1 C 222kPa 5:26
136.4 C 222kPa 5:56
136.2 C 222kPa 6:26
136.2 C 221kPa 6:56
136.3 C 220kPa 7:26
136.4 C 221kPa 7:56
136.4 C 221kPa 7:56
VENTING 7:56
AIR DRYING 8:44
CYCLE COMPLETE 38:44

Digital Signature #
A107F2D2FC8C83FE
