
Midmark M11 - v1.0.5
Total Cycles: 479

Name: RITTER2
Sterilizer ID: V2313228
10/25/2021 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.9 F 0.0
2:00 105.2 F 0.0
4:00 116.6 F 1.8
6:00 146.6 F 5.0
8:00 175.1 F 8.9
10:00 194.9 F 11.6
12:00 212.6 F 7.1
14:00 227.3 F 11.9
16:00 231.8 F 8.8
18:00 239.4 F 11.0
20:00 244.9 F 13.2
22:00 256.7 F 19.6
24:00 267.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.5 F 29.6
2:00 271.9 F 29.5
3:00 272.3 F 29.4
4:00 271.9 F 29.6

Min 271.5 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.3 F 0.0
5:00 199.5 F 0.0
10:00 206.0 F 0.0
15:00 201.6 F 0.0
20:00 195.0 F 0.0
25:00 192.9 F 0.0
30:00 193.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 25:21
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:21
CYCLE COMPLETE

10/25/2021 12:03 PM
ACCEPTED: X
REJECTED:
Operator: U01