
Midmark M11 - v1.0.5
Total Cycles: 1651

Name: RITTER2
Sterilizer ID: V2313228
10/25/2022 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 177.3 F 2.9
4:00 188.6 F 7.5
6:00 211.8 F 12.8
8:00 225.8 F 8.2
10:00 239.4 F 13.0
12:00 239.5 F 11.3
14:00 237.8 F 10.2
16:00 252.5 F 17.0
18:00 264.1 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.4 F 29.7
2:00 271.9 F 29.1
3:00 271.5 F 29.4
4:00 272.9 F 29.5

Min 271.1 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 229.7 F 0.0
10:00 216.4 F 0.0
15:00 225.0 F 0.0
20:00 232.0 F 0.0
25:00 230.0 F 0.0
30:00 214.6 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 19:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:03
CYCLE COMPLETE

10/25/2022 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____