

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000918  
9:21 10/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:17  
PRESSURIZING 3:23  
STERILIZING 4:52  
136.4 C 230kPa 4:52  
136.4 C 232kPa 5:22  
136.0 C 229kPa 5:52  
136.1 C 230kPa 6:22  
136.0 C 228kPa 6:52  
136.3 C 231kPa 7:22  
136.2 C 230kPa 7:52  
136.2 C 230kPa 8:22  
136.2 C 230kPa 8:22  
VENTING 8:22  
AIR DRYING 9:17  
CYCLE COMPLETE 24:17

-----  
Digital Signature #  
094CA7947DBB689C  
-----