
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001655
9:56 31/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:52
STERILIZING 4:54
136.0 C 220kPa 4:54
136.4 C 225kPa 5:24
136.4 C 224kPa 5:54
136.4 C 223kPa 6:24
136.5 C 224kPa 6:54
136.1 C 222kPa 7:24
136.4 C 223kPa 7:54
136.4 C 225kPa 8:24
136.4 C 225kPa 8:24
VENTING 8:24
AIR DRYING 9:16
CYCLE COMPLETE 9:23

Digital Signature #
6775F455AB6EB56A
