
Midmark M11 - v1.0.0
Total Cycles: 2548

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2022 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.1 F 0.0
2:00 121.5 F 0.0
4:00 140.7 F 1.2
6:00 153.2 F 4.0
8:00 175.0 F 7.2
10:00 192.3 F 10.9
12:00 222.3 F 5.1
14:00 218.7 F 7.0
16:00 228.4 F 10.8
18:00 232.2 F 7.8
20:00 240.4 F 12.1
22:00 239.5 F 10.1
24:00 251.0 F 15.7
26:00 261.1 F 21.5
28:00 270.5 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.9
1:00 272.4 F 29.5
2:00 272.4 F 29.2
3:00 271.9 F 29.2
4:00 272.4 F 29.5

Min 271.6 F 28.8
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 207.5 F 0.0
10:00 220.2 F 0.0
15:00 217.6 F 0.0
20:00 221.3 F 0.0
25:00 218.7 F 0.0
30:00 231.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:39
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:08:08
CYCLE COMPLETE

11/11/2022 05:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____